

# Resource Summary Report

Generated by [dkNET](#) on Aug 6, 2026

## Bruker: timsTOF fleX with MALDI-2

RRID:SCR\_023615

Type: Tool

---

### Proper Citation

Bruker: timsTOF fleX with MALDI-2 (RRID:SCR\_023615)

---

### Resource Information

**URL:** <https://www.bruker.com/en/products-and-solutions/mass-spectrometry/timstof/timstof-flex-maldi-2.html>

**Proper Citation:** Bruker: timsTOF fleX with MALDI-2 (RRID:SCR\_023615)

**Description:** Instrument for probing molecular content of sample with MALDI-2. The timsTOF fleX is fully functional high speed, high sensitivity ESI instrument for all X-Omics analyses, combined with high spatial resolution Matrix-Assisted Laser Desorption Ionization (MALDI) source and stage specifically designed for resolving molecular distributions and bringing spatial dimension to OMICS analyses. MALDI-2 uses laser post-ionization to enhance and enrich the MALDI experiment providing access to chemical classes typically opaque to MALDI with enhanced sensitivities.

**Resource Type:** instrument resource

**Keywords:** Thermal Ionization Mass Spectrometry, probing molecular content, X-Omics analyses, instrument, equipment, USEDit

**Availability:** Commercially available

**Specification URL:** [https://raw.githubusercontent.com/SciCrunch/RRID-Instruments/refs/heads/main/PDF/SCR\\_023615.pdf](https://raw.githubusercontent.com/SciCrunch/RRID-Instruments/refs/heads/main/PDF/SCR_023615.pdf)

**Resource Name:** Bruker: timsTOF fleX with MALDI-2

**Resource ID:** SCR\_023615

**Record Creation Time:** 20230527T050216+0000

**Record Last Update:** 20260801T190821+0000

---

### Ratings and Alerts

No rating or validation information has been found for Bruker: timsTOF fleX with MALDI-2.

No alerts have been found for Bruker: timsTOF fleX with MALDI-2.

---

## Data and Source Information

**Source:** [SciCrunch Registry](#)

---

## Usage and Citation Metrics

We have not found any literature mentions for this resource.