

Resource Summary Report

Generated by [dkNET](#) on Aug 5, 2026

Agilent: NovoCyte 3000 Flow Cytometer

RRID:SCR_025890

Type: Tool

Proper Citation

Agilent: NovoCyte 3000 Flow Cytometer (RRID:SCR_025890)

Resource Information

URL: <https://www.agilent.com/en/product/research-flow-cytometry/flow-cytometers/flow-cytometer-systems/novocyte-flow-cytometer-systems-1-3-lasers-984681>

Proper Citation: Agilent: NovoCyte 3000 Flow Cytometer (RRID:SCR_025890)

Description: Agilent NovoCyte 3000 Flow Cytometer with 3 lasers (488, 405, 640 nm) and up 13 colors. 6 colors from violet laser; 5 colors from blue laser; 2 colors from red laser. NovoExpress software used for acquisition.

Synonyms: , Novocyte 3000 Flow Cytometer, Novocyte

Resource Type: instrument resource

Keywords: Agilent, Flow Cytometer, 3 lasers, 13 colors,

Specification URL: https://raw.githubusercontent.com/SciCrunch/RRID-Instruments/main/PDF/SCR_025890.pdf

Resource Name: Agilent: NovoCyte 3000 Flow Cytometer

Resource ID: SCR_025890

Alternate IDs: Model_Number_NovoCyte_3000

Alternate URLs: <https://www.agilent.com/cs/library/datasheets/public/datasheet-technical-specifications-novocyte-5994-1032en-agilent.pdf>,
[https://www.agilent.com/cs/library/usermanuals/public/150014-Novocyte%20Flow%20Cytometer%20Operator's%20Guide%20\(RUO\).pdf](https://www.agilent.com/cs/library/usermanuals/public/150014-Novocyte%20Flow%20Cytometer%20Operator's%20Guide%20(RUO).pdf)

Record Creation Time: 20241016T053254+0000

Record Last Update: 20260801T191340+0000

Ratings and Alerts

No rating or validation information has been found for Agilent: NovoCyte 3000 Flow Cytometer.

No alerts have been found for Agilent: NovoCyte 3000 Flow Cytometer.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We have not found any literature mentions for this resource.